FirstTemp* Genius*

Infrared Thermometer

Caring and Cleaning Guide

Follow these simple instructions to ensure maximum performance from your GENIUS thermometer

- Always ensure the GENIUS thermometer is clean, especially the probe tip lens.
- A dirty lens will reduce the temperatures recorded.

- DO NOT use alcohol on any of the **CLEAR** plastic parts.

- With the lens pointing down gently clean with the applicator. Follow with a lint-free cloth to remove the alcohol film. DO NOT use excessive force.

- To clean the base unit/main body of the GENIUS thermometer (white and blue parts only), use warm water and mild detergent or a less than 10% bleach solution.

- The lens must **always** be clean.
- Wipe using a lint-free cloth.

- Return the thermometer to its base unit and allow 45 minutes drying time prior to using the thermometer.

- The clear plastic display window and base door window should be wiped with a cloth dampened with mild detergent or a less than 10% bleach solution.

If the lens remains soiled:

- Slightly dampen a Q-Tips applicator with Isopropanol.

ALWAYS REMEMBER

1. **Always use a new probe cover.**
2. **Keep probe lens clean.**
3. **Remove the GENIUS thermometer from its base with care and in the proper manner.**

Cleaning Guide

Cleaning Agent	Probe Lens	Clear Plastic Windows	White/Blue Plastic Parts
Mild Detergent	NO	YES	YES
Bleach/Hypochlorite (10%)	NO	YES	YES
Isopropanol	YES	NO	YES
Phenolics	NO	NO	YES

The above cleaning agents are the only ones that should be used with the **GENIUS** thermometer, and must be used in accordance to the **GENIUS** operating manual.

Note:

If the **GENIUS** thermometer has a probe cover installed but is left unused for more than approximately 30 seconds, it enters a power saving SLEEP mode. Installing a **new** probe cover will allow the next temperature to be taken.

* Trademark of Sherwood Services AG © 2004 Tyco Healthcare Group LP. All rights reserved.

P-TH-P-FTGenius01/GB

tyco Healthcare

Tyco Healthcare AG
154 Fareham Road
GOSPORT
HAMPSHIRE
UK - PO13 0AS
Tel: +44 (0) 13 29 22 4258
Fax: +44 (0) 13 29 22 4086

www.tycohealth-ece.com

03/2004